

# Preparation of nitrous oxide ( $\text{N}_2\text{O}$ ) in air standard being traceable to SI by gravimetric method


15th WMO Meeting of Experts on Carbon Dioxide,  
Other Greenhouse Gases, and Related Tracer Measurement Techniques  
**September 7<sup>th</sup>, 2009**  
**Jena, Germany**

Nobuyuki Aoki, Takuya Shimosaka, Nobuhiro Matsumoto, Kenji Kato  
(National Metrology Institute of Japan, AIST)

Keiichi Katsumata, Toshinobu Machida  
(National Institute for Environmental Studies)

# N<sub>2</sub>O standard

WMO Round-Robin intercomparison (2002-2007)



courtesy of Dr. Zhou Lingxi

International comparison among National Metrology Institutes (NMI)

CIPM

International Committee of Weights and Measures

CCQM


the Consultative Committee for Amount of Substance

CCQM-K68

Analysis of N<sub>2</sub>O standard prepared in pilot laboratory (KRISS)

- Component: N<sub>2</sub>O in air
- Concentration: 320nmol/mol
- Time of measurement: 2008-2009
- Status: in progress


# This study



# Gravimetric method traceable to SI

Gravimetric value

$$x_i = \frac{\sum_{A=1}^P \left( \frac{x_{i,A} \cdot m_A}{\sum_{j=1}^n x_{j,A} \cdot M_j} \right)}{\sum_{A=1}^P \left( \frac{m_A}{\sum_{j=1}^n x_{j,A} \cdot M_j} \right)}$$



# Purity analysis of raw materials

Purity table of N<sub>2</sub>O

Components	Applied concentration (μmol/mol)	Standard uncertainty (μmol/mol)	Analytical method
N <sub>2</sub>	0.983	0.567	Micro GC
O <sub>2</sub>	1.020	0.589	Micro GC
CO <sub>2</sub>	0.128	0.006	GC-FID with methanizer
CH <sub>4</sub>	0.019	0.003	GC-FID
CO	0.056	0.032	GC-FID with methanizer
H <sub>2</sub> O	0.439	0.253	Capacitance-type moisture meter
<b>N<sub>2</sub>O</b>	<b>999997.36</b>		

Purity table of N<sub>2</sub>

Components	Applied concentration (μmol/mol)	Standard uncertainty (μmol/mol)	Analytical method
O <sub>2</sub>	0.819	0.473	Micro GC
CO <sub>2</sub>	0.010	0.006	FT-IR
CH <sub>4</sub>	0.026	0.015	FT-IR
CO	0.175	0.101	FT-IR
H <sub>2</sub> O	0.439	0.253	Capacitance-type moisture meter
N <sub>2</sub> O	0.00002	0.00001	Clyo. conc. /GC/MS
<b>N<sub>2</sub></b>	<b>999998.53</b>		


Purity table of O<sub>2</sub>

Components	Applied concentration (μmol/mol)	Standard uncertainty (μmol/mol)	Analytical method
N <sub>2</sub>	0.090	0.052	GC-TCD
Ar	0.087	0.050	GC-TCD
CO <sub>2</sub>	0.056	0.004	FT-IR
CH <sub>4</sub>	0.003	0.002	FT-IR
CO	0.005	0.003	FT-IR
H <sub>2</sub> O	0.439	0.253	Capacitance-type moisture meter
N <sub>2</sub> O	0.00002	0.00001	Clyo. conc. /GC/MS
<b>O<sub>2</sub></b>	<b>999999.32</b>		

Purities of all raw materials are more than **99.999 %**


# N<sub>2</sub>O concentrations in O<sub>2</sub> and N<sub>2</sub>

## Analytical system for impurity N<sub>2</sub>O




	Module 1	Module 2	Module 3
adsorbent	Tenax	Glass beads	Capillary column
Trap Temp.	-120	-150	-180
Desorb Temp	20	150	70

## Calibration standards



N<sub>2</sub>O in He  
(gravimetric method)  
0.149, 1.141, 1.521 nmol/mol



N<sub>2</sub>O concentrations in O<sub>2</sub> and N<sub>2</sub> are less than 0.04 nmol/mol

# Weighing

## Main sources of uncertainty

- Balance
- Buoyancy effect
- Absorption and adsorption to external cylinder surface
- .... etc.

### Comparative Weighing

Mixture




Reference



Alternately-measured

### Weighing room

25° C, 50%




12.123456 kg



### Filling room


24° C, 20~80%



N<sub>2</sub>O or N<sub>2</sub>O in N<sub>2</sub>  
N<sub>2</sub> or O<sub>2</sub>

an hour after filling parent gases  
five hours after filling dilution gases


# Weighing system of NMIJ



Standard deviation of balance system : 2.6 mg !





# Preparation sequence




Five standards of 4th step were independently prepared !

# Verification of intermediated gases





## Internal consistency test

- check that gravimetric value is consistent with analytical value
- highlight significant errors in the preparation process




# N<sub>2</sub>O standards of NMIJ




Cylinder number	Gravimetric Value (nmol/mol)	Expanded uncertainty [k=2] (nmol/mol)	Concentration with calibration curve (nmol/mol)	Deviation (nmol/mol)
CPB00877	303.57	0.13	303.54	0.03
CPB31362	312.77	0.19	312.84	-0.07
CPB16468	326.44	0.20	326.37	0.07
CPC00415	338.20	0.20	338.10	0.10
CPB16463	343.73	0.23	343.84	-0.11

# Comparison between NMIJ and NIES standards



NIES scale is comparable to SI

# Draft report of CCQM-K68



Analysis result of NMIJ is in excellent agreement with key comparison reference value (KCRV)

# Summary

- N<sub>2</sub>O standards **traceable to SI** have been prepared at NMIJ
- The scale of NIES-standards is **linkage to the scale of NMIJ-standards**
- Analysis result of NMIJ is in excellent agreement with KCRV in CCQM-K68
- The scale of NIES-standards will be **comparable to NMI**



*National Institute of  
Advanced Industrial Science  
and Technology*

**AIST**

# International comparison among NMi

- **Measurement capability**

CCQM-K68 : Nitrous oxide 320 nmol/mol in artificial Air

- **Gravimetric preparation of N<sub>2</sub>O by coordinating lab.**

Purity Assessment + Weighing Technique + Stability + Verification

N<sub>2</sub>O 300 ~ 350 nmol/mol  $\pm$  0.06 % (95 % confidence level)

- **How far does the light shine**

This key comparison will support the measurement capability of N<sub>2</sub>O at ambient level

- **Participation(8)**

KRISS, NIM, NIST, NMIJ, GMD/NOAA, VSL, IMK-IFU, VNIIM



# Uncertainty

## Main source of uncertainty

- Purity of raw material
- Balance
- mass piece
- adsorption/desorption
- buoyancy effect
- molar mass